

BASIC OPERATIONAL REQUIREMENTS FOR MAMMOGRAPHY

- Registration of machine
- FDA certification of Facility
- ACR accreditation of Machine

And in addition:

- Compliance with all NAC 459 general radiation safety regulations
- Compliance with NAC 457.300 QA program in compliance with 21 CFR part 900
- Physician supervision under NAC 457.305 and .340
- Retention of machine records
- Issuance of reports in compliance with NAC 457.313 in 7 days for preliminary; 30 days final unless it is a severe finding, and lay report in 30 days.
- Qualified physician readers NAC 457.345
- Qualified Techs NAC 457.350